

UK Declaration of Conformity

Apparatus:

Mag613

Description:

An externally DC powered fluxgate magnetic field sensor, that connects to other equipment, and is used for the measurement of static and alternating magnetic fields in three axes.

The object of the declaration described above is in conformity with the relevant regulations:

The Electromagnetic Compatibility Regulations 2016 UK SI 2016 No. 1091
The Restriction of the Use of Certain Hazardous Substances in Electrical and Electronic Equipment Regulations 2012 UK SI 2012 No. 3032

This declaration of conformity is issued under the sole responsibility of the manufacturer.

Test Specifications:

The above product conforms to the following specifications:

BS EN 61326-1: 2021 - Electrical equipment for measurement, control and laboratory use

Related Documents:

EN 55011 Radiated Emissions – E-Field (Class B)
EN 61000-4-2 Immunity to Electrostatic Discharge
EN 61000-4-3 Immunity to Radiated Fields
EN 61000-4-6 Immunity to Conducted Disturbances

This is to certify that the apparatus described above conforms with the protection requirements of BS EN 61326-1: 2021.

Signed: Detrik

Daniel Cetnik
Technical Director
Bartington Instruments Ltd

Date of Issue: 08/04/2024
Place of Issue: Witney, Oxfordshire, UK

DC1854-UK Issue 1

DCN No: 2474

Telephone: +44 (0)1993 706565 • **Email:** sales@bartington.com • **Website:** www.bartington.com